



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 16-Apr-2014

MY03S11143154-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8, KANJANAVANICH ROAD
TAMBOL PRIK, AMPHUR SADAO
SONGKHLA

THAILAND

Order Information

Account Number:	101195
Synergy Health Sales Part Reference:	1024139
Customer Reference Number:	4027005540
Product Description:	KIMTECH Pure* G3 Sterile Sterling Nitrile Gloves, 12" Hand Specific Pairs
Validation Reference:	0.0358
Quantity Received:	792
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	50.0
Customer Unit Lot/Batch Number:	SEE BELOW



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 16-Apr-2014

MY03S11143154-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
 200 MOO 8, KANJANAVANICH ROAD
 TAMBOL PRIK, AMPHUR SADAO
 SONGKHLA

THAILAND

Other Process Details:

Catalog Number	Batch No/Lot No.	Quantity
11822	970414/SM40932XX	36
11822	970414/SM40932XX	7
11822	970414/SM40932XX	29
11822	970414/SM40932XX	36
11823	970414/SM40912XX	36
11823	970414/SM40922XX	9
11823	970414/SM40932XX	27
11823	970414/SM40922XX	36
11823	970414/SM40922XX	14
11823	970414/SM40922XX	22
11823	970414/SM40932XX	17
11823	970414/SM40952XX	19
11824	970414/SM40942XX	9
11824	970414/SM40952XX	27
11824	970414/SM40942XX	36
11825	970414/SM40912XX	36
11825	970414/SM40912XX	3
11825	970414/SM40922XX	33
11825	970414/SM40912XX	36
11825	970414/SM40922XX	36
11825	970414/SM40922XX	4
11825	970414/SM40922XX	32
11825	970414/SM40922XX	6
11825	970414/SM40952XX	22
11825	970414/SM40952XX	36
11825	970414/SM40952XX	8
11825	970414/SM40952XX	36



www.synergyhealthpic.com

Certificate of Irradiation

Date Issued: 16-Apr-2014

MY03S11143154-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
 200 MOO 8, KANJANAVANICH ROAD
 TAMBOL PRIK, AMPHUR SADAO
 SONGKHLA

THAILAND

11822	970414/SM40932XX	26
11822	970414/SM40972XX	10
11822	970414/SM40972XX	36
11822	970414/SM40972XX	6
11822	970414/SM40982XX	30
11823	970414/SM40952XX	29
11823	970414/SM40962XX	7

Irradiation Data

Date and Time of Irradiation:	15-Apr-2014 03:03
Reference Dose Range kGy:	30.8 - 31.4
Calculated Minimum Dose kGy:	25.0
Calculated Maximum Dose kGy:	34.6

NORAIZA ZAKARIA
 QA Executive/Senior Chemist
 Synergy Sterilisation (M) Sdn. Bhd.
 +60(0)44152111

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number:

Page 3 of 3