

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 03-Jul-2014 UK33S11190875-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Order Information

Account Number

Synergy Health Sales Part Reference:

Customer Reference Number:

Product Description:

Validation Reference:

Quantity Received:

Customer Minimum Specification kGy:

Customer Maximum Specification kGy:

Customer Unit Lot/Batch Number:

Other Process Details:

124123

1073384

0000006492

POLYCHLOR GLOVES GP6 DV4162 25-45kGy

4162

48

25.0

45.0

0000006492

Irradiation No: S1119087531

Description: B/Fusion St PCP Glove (S-BFAP)

Packing: 200 Pairs/ ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-0640 C31404004 S-BFAP-7075 25

Description: Bioclean P.Zero St PCP (BPZS)

Packing: 200 Pairs/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-0758 C21404023 BPZS75 23

Total: 48 Cartons

Irradiation Data

Date and Time of Irradiation: 03-Jul-2014 05:08



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 03-Jul-2014 UK33S11190875-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Reference Dose Range kGy:	33.3 - 34.0
Calculated Minimum Dose kGy:	29.4
Calculated Maximum Dose kGy:	36.1