

Certificate No: **13010019**

## CERTIFICATE OF IRRADIATION

**CUSTOMER** : NITRITEX (M) SDN BHD

**ADDRESS** : No:2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,  
40150 Shah Alam, Selangor Darul Ehsan.

**DOSE REQUESTED** : MINIMUM : 25 kGy MAXIMUM : 38 kGy

**IRRADIATION DATE** : 02.01.2013 **IRRADIATION LOT NO.** : 13010019

**PRODUCT DESCRIPTION** : Bioclean D St. Cvrl Hood (S-BDCHT)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 20 Pieces/ctn			
12-1847	-	S-BDCHT-S	18
12-1848	-	S-BDCHT-M	36
<b>Total</b>			<b>54</b>

has been successfully irradiated and received the dose specification requested (as the above stated)  
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 27.9 kGy Max. Dose Delivered (kGy) : 31.8 kGy

Approved By

*Noraini* 02/01/2013  
Noraini binti Sudin  
Senior Executive QA  
Synergy Sterilisation Rawang (M) Sdn. Bhd.  
(234258-U)