



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 01-Jul-2015

UK33S11409773-9-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
 EN ISO 9001:2008 Quality Management System  
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd  
 Minton Enterprise Park  
 Oaks Drive  
 Newmarket  
 Suffolk  
 CB8 7YY

## Order Information

Account Number:	124123																				
Synergy Health Sales Part Reference:	1073641																				
Customer Reference Number:	0000007195																				
Product Description:	BIOCLEAN C&D GROUP 27 DV4166 25-38kGy																				
Validation Reference:	4166																				
Quantity Received:	41																				
Customer Minimum Specification kGy:	25.0																				
Customer Maximum Specification kGy:	38.0																				
Customer Unit Lot/Batch Number:	0000007195																				
Other Process Details:	Irradiation No: S1140977391 Description: Bioclean D St. Cvrl Collar (S-BDCCT) Packing: 180 Pieces/ ctn																				
	<table border="1"> <thead> <tr> <th>Lot No</th> <th>Batch No</th> <th>Product Code / Size</th> <th>Qty(ctns)</th> </tr> </thead> <tbody> <tr> <td>15-0514</td> <td>N/A</td> <td>S-BDCCT-L</td> <td>5</td> </tr> <tr> <td>15-0994</td> <td>N/A</td> <td>S-BDCCT-L</td> <td>24</td> </tr> <tr> <td>15-0213</td> <td>N/A</td> <td>S-BDCCT-S</td> <td>6</td> </tr> <tr> <td>15-0993</td> <td>N/A</td> <td>S-BDCCT-S</td> <td>6</td> </tr> </tbody> </table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	15-0514	N/A	S-BDCCT-L	5	15-0994	N/A	S-BDCCT-L	24	15-0213	N/A	S-BDCCT-S	6	15-0993	N/A	S-BDCCT-S	6
Lot No	Batch No	Product Code / Size	Qty(ctns)																		
15-0514	N/A	S-BDCCT-L	5																		
15-0994	N/A	S-BDCCT-L	24																		
15-0213	N/A	S-BDCCT-S	6																		
15-0993	N/A	S-BDCCT-S	6																		
	Total: 41 Cartons																				

## Irradiation Data

Date and Time of Irradiation:	01-Jul-2015 06:09
Reference Dose Range kGy:	33.1 - 34.4
Calculated Minimum Dose kGy:	30.3
Calculated Maximum Dose kGy:	36.7



<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 01-Jul-2015

UK33S11409773-9-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
EN ISO 9001:2008 Quality Management System  
EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd  
Minton Enterprise Park  
Oaks Drive  
Newmarket  
Suffolk  
CB8 7YY

Irradiation Release Authorised By Synergy Health plc