



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 14-Mar-2014

MY01S11127107-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process

in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD

No.2, Jalan Jurunilai U1/20

Seksyen U1, Hicom Glenmarie

Industrial Park

40150 Shah Alam

Selangor

Order Information

Account Number:	101178								
Synergy Health Sales Part Reference:	1055050								
Customer Reference Number:	SNB14/0063								
Product Description:	Bioclean DISPOSABLE								
Validation Reference:	0.0098								
Quantity Received:	36								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	38.0								
Customer Unit Lot/Batch Number:	SNB14/0063								
Other Process Details:	Irradiation No: S1112710711 Description: Bioclean D St. Overboot (S-BDOB) Packing: 150 Pieces/ ctn								
	<table border="1"> <thead> <tr> <th>Lot No</th> <th>Batch No</th> <th>Product Code / Size</th> <th>Qty(ctns)</th> </tr> </thead> <tbody> <tr> <td>14-0154</td> <td>-</td> <td>S-BDOB</td> <td>36</td> </tr> </tbody> </table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	14-0154	-	S-BDOB	36
Lot No	Batch No	Product Code / Size	Qty(ctns)						
14-0154	-	S-BDOB	36						
	Total : 36 Cartons								

Irradiation Data

Date and Time of Irradiation:	14-Mar-2014 09:56
Reference Dose Range kGy:	34.4 - 35.2
Calculated Minimum Dose kGy:	31.2
Calculated Maximum Dose kGy:	37.4

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01, 18th Floor, MWE Plaza, No. 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number:

Page 1 of 1