

CERTIFICATE OF CONFORMITY

PRODUCT: Sterile Isopropanol 70% v/v with WFI Quality Water

PRODUCT CODE: IPAWF11-5LS

PRODUCT DESCRIPTION: Low Residue, Rapid Action, Fast Drying, Alcohol Disinfectant

BATCH NUMBER: 029955

MANUFACTURING DATE: Dec 2013

EXPIRY DATE: Nov 2015

IRRADIATION NUMBER: S11086744-1-1

IRRADIATION DOSE: 25kGy's to 45 kGy's

STERILITY TEST REF: To Follow

CERTIFICATE OF ANALYSIS No: 029955

PRODUCT TEST RESULT: Certified that the goods listed above have been inspected, and unless otherwise stated conform in all respects to the requirements of the product specification.

APPROVAL	
APPROVED BY	<i>Keith Hanon</i>
POSITION	FACILITY OPERATOR
DATE	8TH JAN 14

CERTIFICATE OF ANALYSIS

PRODUCT: Sterile Isopropanol 70% v/v with WFI Quality Water

PRODUCT CODE: IPAWF11-5LS

PRODUCT DESCRIPTION: Sterile Isopropanol Alcohol in WFI quality water

BATCH NUMBER: 029955

MANUFACTURING DATE: Dec 2013

EXPIRY DATE: Nov 2015

PRODUCT ASSESSMENT		
FILTRATION:	0.2 micron in line filter	PASS
COLOUR:	Colourless Liquid	Pass
ODOUR:	Typical	Pass
CLARITY:	Clear	Pass

APPROVAL	
APPROVED BY	<i>Kevin Hanson</i>
POSITION	FACILITY OPERATOR
DATE	8TH JAN 14



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 13-Dec-2013

UK32S11086744-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

- EN ISO 11137-1:2006 Sterilisation of Health Care Products
- EN ISO 9001:2008 Quality Management System
- EN ISO 13485:2012 Quality System - Medical Devices

Redditch Medical
Discovery 2
2 William Armstrong Way
Net Park
Sedgefield
Co.Durham TS21 3FH

Order Information

Account Number:	101254
Synergy Health Sales Part Reference:	1006677
Customer Reference Number:	8211
Product Description:	INSPEC DE 5 LTR BOTTLES 25-45KG
Validation Reference:	4.1373
Quantity Received:	102
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	2 PLTS; THUR 12.12.13; 101 x B/NO. 029955 + 1 SAMPLE

Irradiation Data

Date and Time of Irradiation:	13-Dec-2013 00:06
Reference Dose Range kGy:	38.7 - 42.3
Calculated Minimum Dose kGy:	27.0
Calculated Maximum Dose kGy:	42.3

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069