

CERTIFICATE OF CONFORMITY

PRODUCT: InSpec OX 5L Bottle

PRODUCT CODE: OXPW40-5LS

PRODUCT DESCRIPTION: Sterile Peracetic Acid / Hydrogen Peroxide

BATCH NUMBER: 027574

MANUFACTURING DATE: Jun 2013

EXPIRY DATE: May 2014

IRRADIATION NUMBER: S11014288-1-1

IRRADIATION DOSE: 25kGy's to 45 kGy's

CERTIFICATE OF ANALYSIS No: 027574

PRODUCT TEST RESULT: Certified that the goods listed above have been inspected, and unless otherwise stated conform in all respects to the requirements of the product specification.

APPROVAL	
APPROVED BY	<i>Keith Hancock</i>
POSITION	<i>FACILITY OPERATOR</i>
DATE	<i>17TH JUNE 13</i>

CERTIFICATE OF ANALYSIS

PRODUCT: InSpec OX 5L Bottle

PRODUCT CODE: OXPW40-5LS

PRODUCT DESCRIPTION: Sterile Peracetic Acid / Hydrogen Peroxide

BATCH NUMBER: 027574

MANUFACTURING DATE: Jun 2013

EXPIRY DATE: May 2014

PRODUCT ASSESSMENT		
FILTRATION:	0.2 micron in line filter	PASS
COLOUR:	Colourless Liquid	Pass
ODOUR:	Typical	Pass
CLARITY:	Clear	Pass

APPROVAL	
APPROVED BY	<i>Keith Hansen</i>
POSITION	FACILITY OPERATOR
DATE	17 TH JUNE 13



http://www.synergyhealth.co.uk

Certificate of Irradiation

Date Issued: 23-Apr-2013

UK32S11014288-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

ISO 11137:2006 Sterilisation of Health Care Products
ISO 9001:2008 Quality Management System
ISO 13485:2012 Quality System - Medical Devices

Redditch Medical
Discovery 2
2 William Armstrong Way
Net Park
Sedgefield
Co.Durham TS21 3FH
UNITED KINGDOM

Order Information

Customer Reference Number:	7652
Product Description:	EMPTY BOTTLES 25-45KGY
Quantity Received:	156
Customer Unit Lot/Batch Number:	8 PLTS LOT 2 TUES 2.4.13
Other Process Details:	113 X OXPW40-5LS , 10 X ANWF120 - 1LS 900ML , 12 X ANWF120-5LS , 10 X QTWF130-5LS , 10 X QTWF130-1LS 900ML. Validation 4.1448 carried out on this consignment. Min dose = 29.8 kGy, Max dose = 36.7 kGy.
Shop Order Number:	S11014288-1-1
Synergy Health Sales Part Reference:	1015894

Irradiation Data

Date and Time of Irradiation:	20-Apr-2013 22:53
-------------------------------	-------------------

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069