



525 Locust Grove
Spartanburg, SC 29303

864-503-8333
1-800-289-5762

fax 864-503-8444
www.contecinc.com

CERTIFICATE OF CONFORMANCE AND STERILITY

Date: March 12, 2013

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Product Code: PSME0001
Lot Number: 271616-*-*-1
Manufacturing Date: 02/28/2013
Isopropyl Alcohol: 70 +/- 3% by volume with the remainder DI H2O.
Run Number: 1049120
Irradiation Dose Specified: 25kGy to 50kGy
Irradiation Dose Delivered: 28.4kGy to 44.3kGy
Validated Sterile Product: Validation Sterile packets available upon request.
Sterility Assurance Level: 10^{-6} (SAL determined by test method
AAM1/ISO 11137)
Expiration Date: 02/2015

QC Supervisor
coc@contecinc.com



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102
 STER0002

03/12/13 15:21:20 GMT
 Page - 1 of 1

Customer Name: Contec-Box 530
 P.O.# 58060

Processing Facility: Charlotte Gamma

Work Order # 1049120
 Sales Order # 920179

25-50 kGy

CNT1, Gamma Treatment

Irradiation Date/Time: 03/10/13 14:32:00 GMT

SO Line #	Qty	UOM	Customer Item Number / Customer Item Description	Customer Lot Number / Customer Load Number
101.00	48	CA	6-LS7030VS-PC Clean Room Product	266782-**-1 NA
102.00	32	CA	6-LS7030VS-PC Clean Room Product	266784-**-1 NA
103.00	36	CA	PSC20005 Clean Room Product	269639-**-1 NA
104.00	30	CA	PSME0001 Clean Room Product	271616-**-1 NA
105.00	24	CA	PSPF0039 Clean Room Product	269642-**-1 NA
106.00	54	CA	PSPS0078 Clean Room Product	266808-**-1 NA
107.00	54	CA	PSPS0093 Clean Room Product	266843-**-1 NA
108.00	128	CA	PSW5V3797 Clean Room Product	260887-**-1 NA
	406	CA	Total	

Quality Test Summary

-----Signed By-----

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date / Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	28.4 kGy	Pass	CCAMPS Charlie Camps	03/11/13 01:26:53 GMT
				Reason Code Test			
450.00	Maximum Dose	25.0 kGy	50.0 kGy	44.3 kGy	Pass	CCAMPS Charlie Camps	03/11/13 01:27:21 GMT
				Reason Code Test			

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ALEX HERRINGTON
 Work Order Completions

Date: 03/12/13 15:18:50 GMT