



Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA.
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306
Website: <https://www.synergyhealthplc.com>

Certificate No: **13031262**

CERTIFICATE OF IRRADIATION

CUSTOMER : NITRITEX (M) SDN BHD
ADDRESS : No:2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,
40150 Shah Alam, Selangor Darul Ehsan.
DOSE REQUESTED : MINIMUM : 25 kGy MAXIMUM : 40 kGy
IRRADIATION DATE : 12.03.2013 **IRRADIATION LOT NO.** : 13031262
PRODUCT DESCRIPTION : Bioclean Extra 400 St Latex (BLAS)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 200 Pairs/ctn			
13-0361	L61303001	BLAS-M	25
13-0368	L61303003	BLAS-L	90
13-0314	L61302030	BLAS-S	45
Total			160

has been successfully irradiated and received the dose specification requested (as the above stated)
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 25.3 kGy Max. Dose Delivered (kGy) : 38.3 kGy

Approved By

: *Noraini* 12/02/2013
.....
Noraini binti Sudin
Senior Executive QA
Synergy Sterilisation Rawang (M) Sdn. Bhd.
(234258-U)