



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 13-Feb-2014

MY01S11111291-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
 No.2, Jalan Jurunilai U1/20
 Seksyen U1, Hicom Glenmarie
 Industrial Park
 40150 Shah Alam
 Selangor

Order Information

Account Number:	101178								
Synergy Health Sales Part Reference:	1055042								
Customer Reference Number:	SNB14/0038								
Product Description:	Gloves								
Validation Reference:	11/1(Nitritex)(03)								
Quantity Received:	54								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	40.0								
Customer Unit Lot/Batch Number:	SNB14/0038								
Other Process Details:	Irradiation No: S1111129111 Description: Bioclean Alpha St Latex (AL300) Packing: 200 Pairs/ ctn								
	<table border="1"> <thead> <tr> <th>Lot No</th> <th>Batch No</th> <th>Product Code / Size</th> <th>Qty(ctns)</th> </tr> </thead> <tbody> <tr> <td>14-0171</td> <td>L31401037</td> <td>AL300-80</td> <td>54</td> </tr> </tbody> </table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	14-0171	L31401037	AL300-80	54
Lot No	Batch No	Product Code / Size	Qty(ctns)						
14-0171	L31401037	AL300-80	54						
	Total : 54 Cartons								

Irradiation Data

Date and Time of Irradiation:	12-Feb-2014 20:00
Reference Dose Range kGy:	26.5 - 31.7
Calculated Minimum Dose kGy:	26.1
Calculated Maximum Dose kGy:	32.3

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01, 18th Floor, MWE Plaza, No. 8, Lebuh Farquhar, 10200 , Penang , MALAYSIA

VAT Number: