



Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA.
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306
Website: <https://www.synergyhealthplc.com>

Certificate No: **13115954**

CERTIFICATE OF IRRADIATION

CUSTOMER : NITRITEK (M) SDN BHD
ADDRESS : No:2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,
40150 Shah Alam, Selangor Darul Ehsan.
DOSE REQUESTED : MINIMUM : 25 kGy MAXIMUM : 40 kGy
IRRADIATION DATE : 15.11.2013 IRRADIATION LOT NO. : 13115954
PRODUCT DESCRIPTION : Bioclean Suprene St PCP (BSNS)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 200 Pairs/ctn			
13-1761	C61310021	BSNS-70	54
Total			54

has been successfully irradiated and received the dose specification requested (as the above stated)
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 26.8 kGy Max. Dose Delivered (kGy) : 32.7 kGy

Approved By :

f f d . 18/11/13
Noraini binti Sudin
Senior Executive QA
Synergy Sterilisation Rawang (M) Sdn. Bhd.
(234258-U)