

CERTIFICATE OF IRRADIATION / STERILIZATION CERTIFICATE

THIS IS TO CERTIFY THAT


Type of product: DermaShield®
Style: 73-701
Manufacturing Lot No: 1208 5283 04
Requested dose (min): 25 kGray
Manufacturing Date: 2012/07
Expiry Date: 2017/06

HAS BEEN SUCCESSFULLY IRRADIATED & HAS BEEN STERILIZED USING GAMMA RADIATION

Location: Ansell Malaysia Sdn. Bhd, Melaka, Malaysia

Irradiation Lot No.: 1208 5283 04
Irradiation date: 02-10/08/2012
Dose delivered (min): 25 kGray

CERTIFIED BY



Guido Van Duren

*Director Global Regulatory Compliance
PPE Products*

Date: 11/04/2013