



www.synergyhealthplc.com

## Certificate of Irradiation

Date Issued: 24-Feb-2014 – Rev01

MY01S11096760-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process

in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
 EN ISO 9001:2008 Quality Management System  
 EN ISO 13485:2012 Quality System - Medical Devices

NITRITEK (M) SDN BHD  
 No.2, Jalan Jurunilai U1/20  
 Seksyen U1, Hicom Glenmarie  
 Industrial Park  
 40150 Shah Alam  
 Selangor

### Order Information

Account Number:	101178									
Synergy Health Sales Part Reference:	1055039									
Customer Reference Number:	SNB14/0012									
Product Description:	Bioclean St. Mask									
Validation Reference:	0.0028									
Quantity Received:	60									
Customer Minimum Specification kGy:	25.0									
Customer Maximum Specification kGy:	38.0									
Customer Unit Lot/Batch Number:	(MTA210-1)									
Other Process Details:	Irradiation Lot No: S1109676011 Description: Bioclean St. Mask Tie On Indv (MTA210-1) Packing: 600 Pieces/Ctn									
	<table border="1"> <thead> <tr> <th>Lot No</th> <th>Product Code / Size</th> <th>Qty(Ctns)</th> </tr> </thead> <tbody> <tr> <td>0310282</td> <td>MTA210-1</td> <td>32</td> </tr> <tr> <td>0310283</td> <td>MTA210-1</td> <td>28</td> </tr> </tbody> </table>	Lot No	Product Code / Size	Qty(Ctns)	0310282	MTA210-1	32	0310283	MTA210-1	28
Lot No	Product Code / Size	Qty(Ctns)								
0310282	MTA210-1	32								
0310283	MTA210-1	28								
	Total: 60 Cartons									

### Irradiation Data

Date and Time of Irradiation:	14-Jan-2014 07:54
Reference Dose Range kGy:	33.8 - 34.0
Calculated Minimum Dose kGy:	30.4
Calculated Maximum Dose kGy:	34.9

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01, 18th Floor, MWE Plaza, No. 8, Lebuh Farquhar, 10200, Penang, MALAYSIA

VAT Number: