

Synergy Health (Thailand) Ltd.
700/465 Amata Nakorn Industrial
Moo 7 Tambon Donhuaroh
Amphur Muang Chonburi, Chonburi 20000

T. +66 (0)38 458 431 to 3
+66 (0)38 450 092 to 3
F. +66 (0)38 458 435
+66 (0)38 717 146

CERTIFICATE OF IRRADIATION

This is to certify that, under the irradiation batch number

T2014 2388

for the account of **THP 311**

Synergy Health (Thailand) Ltd. treated by gamma radiation from cobalt-60 products that the customer said to be:-

Product name **Kimtech Pure M3 Sterile Mask**
Total quantity **1,812 cartons**
Purchase order no. **P571833**
Product description

<i>Product</i>	<i>Lot no.</i>	<i>Quantity (cartons)</i>
6246701	TH42676946	180
6247001	TH42586946	720
6249401	TH42486946	912

The process was applied according to specification: **T-B311192513**

The first tote was irradiated on: **4 October 2014**

The applied doses were checked with calibrated dosimeters traceable to a national laboratory of metrology.

The dose specification is **25.0** kilograys minimum and **50.0** kilograys maximum

The calculated dose received is **31.1** kilograys minimum and **41.4** kilograys maximum

Sudarat (for) 4 Oct. 14

Approved by/date : **Paradee Thammajanya, QA Manager**