



Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)  
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,  
48000 RAWANG, SELANGOR, MALAYSIA.  
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306  
Website: <https://www.synergyhealthplc.com>

Certificate No: **13020908**

## CERTIFICATE OF IRRADIATION

**CUSTOMER** : NITRITEX (M) SDN BHD  
**ADDRESS** : No.2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,  
40150 Shah Alam, Selangor Darul Ehsan.  
**DOSE REQUESTED** : MINIMUM : 25 kGy MAXIMUM : 38 kGy  
**IRRADIATION DATE** : 23.02.2013 **IRRADIATION LOT NO.** : 13020908  
**PRODUCT DESCRIPTION** : Bioclean St. Mask E.loop Indv (MEA210-1)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 600 pieces / ctn			
01210286	-	MEA210-1	26
01210289	-	MEA210-1	40
01210290	-	MEA210-1	40
01210291	-	MEA210-1	02
<b>Total</b>			<b>108</b>

has been successfully irradiated and received the dose specification requested (as the above stated)  
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 30.0 kGy Max. Dose Delivered (kGy) : 37.4 kGy

Approved By

*Noraini* 25/02/2013

Noraini binti Sudin  
Senior Executive QA  
Synergy Sterilisation Rawang (M) Sdn. Bhd.  
(234258-U)