



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 29-Sep-2014

UK33S11242642-5-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123								
Synergy Health Sales Part Reference:	1073641								
Customer Reference Number:	0000006665								
Product Description:	BIOCLEAN C&D GROUP 27 DV4166 25-38kGy								
Validation Reference:	4166								
Quantity Received:	32								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	38.0								
Customer Unit Lot/Batch Number:	0000006665								
Other Process Details:	Irradiation No: S1124264251 Description: Bioclean C Disp St Sleevecover (S-BCSC) Packing: 180 Pieces/ ctn								
	<table border="1"> <thead> <tr> <th>Lot No</th> <th>Batch No</th> <th>Product Code / Size</th> <th>Qty(ctns)</th> </tr> </thead> <tbody> <tr> <td>14-1451</td> <td>-</td> <td>S-BCSC</td> <td>32</td> </tr> </tbody> </table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	14-1451	-	S-BCSC	32
Lot No	Batch No	Product Code / Size	Qty(ctns)						
14-1451	-	S-BCSC	32						
	Total: 32 Cartons								

Irradiation Data

Date and Time of Irradiation:	25-Sep-2014 14:06
Reference Dose Range kGy:	32.2 - 32.2
Calculated Minimum Dose kGy:	29.5
Calculated Maximum Dose kGy:	34.3

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069