

DOCUMENT NO: QP-07-COI
REVISION NO: 6/02

STERILGAMMA (M) SDN. BHD. (234258-U)

LOT 42, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA
TEL: +60(3) 6091 5301 (4 lines) FAX: +60(3) 6091 5320
Website: <http://www.sterilgamma.com>
email: feedback@sterilgamma.com

Certificate of Irradiation

This is to certify that

Customer: **NITRITEX (M) SDN BHD**
Address: **No: 2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie
Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan.**
Type of Product: **Bioclean Permaflow Sterile Cleanroom Marker Pen**
Manufacturing Lot No: **Refer Attachment No: (4158)**
No. of Cartons: **15 Cartons**
Requested Dose (min): **25.0 kGy**

has been successfully irradiated and received the minimum dose requested (as the above stated) within the precision and accuracy of dosimetry system employed

Irradiation Lot No: **11104158**
Irradiation Date: **16 October, 2011**
Carton Serial No: **836048 - 836062**
Dose Delivered (min): **29.0 kGy** Max: **31.2 kGy**

Certified By



Designation: **Head Of Quality Assurance**

Date: **18 October, 2011**

DOCUMENT NO: QP-07-COI
REVISION NO: 6/02

STERILGAMMA (M) SDN. BHD. (234258-U)

LOT 42, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA
TEL: +60(3) 6091 5301 (4 lines) FAX: +60(3) 6091 5320
Website: <http://www.sterilgamma.com>
email: feedback@sterilgamma.com

Attachment No : (4158)

Irradiation Lot No : 11104158

Page : 1 of 1

Product Code	Mfg Lot No./Batch No.	Size	Unit (Pcs)	No. Of Ctns
Bioclean Permaflow Sterile Cleanroom Marker Pen	11-1156	(S-BPFP-BK)	1,200	6
	11-1157	(S-BPFP-BL)	400	2
	11-1158	(S-BPFP-RD)	400	2
	11-1163	(S-BPTB-BK)	800	4
	11-1164	(S-BPTB-BL)	200	1
Total			3,000	15

Verify by



Production Manager
SterilGamma (M) Sdn Bhd
Date: October 18, 2011